

# **GLOVE SPECIFICATIONS**

Item Number	LA4030
Product Group	Disposable Gloves
Description	<ul> <li>Dura-Cor™, Latex Examination Gloves, Powder Free, 11-mil Palm Thickness, 300 mm. length (12 inches), Blue Color, Beaded Cuff, Textured Finish, Low Protein Content</li> </ul>

## LA4030

# Class 1 Medical Device

**U.S:** Complies with FDA 510(k) requirements. **Europe**: Complies with 93/42/EEC

ASTM D3578: Standard Specifications for Rubber Exam Gloves

ASTM D5151: Standard Test Method for Detection of Holes in Medical Gloves

Accepted Quality Level (AQL): 1.5

**ISO 9001: 2000** Certified Manufacturing Facility

#### **Food Contact Information:**

U.S.: Component Materials comply with all federal regulations for food contact. (U.S. Code of Federal Regulations, 21CFR, Parts 170-199)
 Europe: Complies with 90/128/EEC





## **Mailing Address:**

Eastern Glove & Safety P.O. Box 1484 Havertown, PA 19083



**Shipping Address:** 900 Calcon Hook Road Sharon Hill, PA 19079

sales@easternglove.com www.easternglove.com

Phone: 610-324-4595 Fax: 610-446-1143